

Certificate

Traceability of Manufacturers to the IFCC Reference Measurement Procedure for HbA1c

This certifies that Siemens Healthcare Diagnostics using DCA Vantage, participates in the Monitoring Programme to demonstrate traceability. In the Monitoring Programme of 2014 the following performance was seen:

Deviation from IFCC-target	at 30 mmol HbA1c/mol Hb :	2.1
	at 60 mmol HbA1c/mol Hb :	2.6
	at 90 mmol HbA1c/mol Hb :	3.0
Reproducibility, coefficient of variation		3.39%
Linearity, correlation coefficient		0.9955

Date of issue: 10 December 2014

Certification expires: 31 December 2015



IFCC Network Coordinator